



Customer: Grove, Inc.
 Customer Sample ID: 800mg Watermelon OG #201111WMOG
 Laboratory Number: 20K0366-02



Cannabinoid Profile

Extraction Technician: DF
 Analytical Chemist: CB

| Extraction Date(s) | Analysis Date(s) |
|--------------------|------------------|
| 11/18/2020 | 11/18/2020 |

| Cannabinoids (HPLC) | | Results | |
|---|------------|----------|-------------|
| | LOD (mg/g) | % | mg/g |
| Cannabidivarin (CBDV) | <0.40 | | |
| Cannabidiolic Acid (CBD-A) | <0.40 | | |
| Cannabigerolic Acid (CBG-A) | <0.40 | | |
| Cannabigerol (CBG) | <0.40 | | |
| Cannabidiol (CBD) | <0.40 | | |
| Tetrahydrocannabivarin (THCV) | <0.40 | | |
| Cannabinol (CBN) | <0.40 | | |
| delta 9-Tetrahydrocannabinol (THC) | <0.40 | | |
| delta 8-Tetrahydrocannabidol | | 75.71 | 757 |
| Cannabichromene (CBC) | <0.40 | | |
| delta-9-Tetrahydrocannabinolic Acid (THC-A) | <0.40 | | |
| Cannabinoids Total | | % | mg/g |
| Max Active THC | | 0.00 | 0.00 |
| Max Active CBD | | 0.00 | 0.00 |
| T.Active Cannabinoids | | 0.00 | 0.00 |
| Total Cannabinoids | | 75.70 | 757.00 |

Following USDA guidelines on uncertainty, Altitude Consulting's uncertainty are calculated for CBDa and CBD at +/- 4%. The uncertainty for THCa and THC are +/- 5%. This implies the range for a 10% value of CBD to be 9.6-10.4%. The uncertainty range for a 0.30% value of THC would be 0.28-0.32%. The measurement uncertainty is calculated using a coverage factor of 2.

Cannabinoid (mg/g)



| | | | | |
|---|--|---|---|---|
| ■ Cannabichromene (CBC) | ■ Cannabidiol (CBD) | ■ Cannabidiolic Acid (CBD-A) | ■ Cannabidivarin (CBDV) | ■ Cannabigerol (CBG) |
| ■ Cannabigerolic Acid (CBG-A) | ■ Cannabinol (CBN) | ■ delta 8-Tetrahydrocannabidol | ■ delta 9-Tetrahydrocannabinol (THC) | ■ delta-9-Tetrahydrocannabinolic Acid (THC-A) |
| ■ Tetrahydrocannabivarin (THCV) | | | | |

Reporting Limits will vary based on sample extraction weight used for the analysis.

Altitude Consulting, LLC utilizes NIST traceable Reference Standards and Certified Reference Material to calibrate analytical instruments along with proven analytical methods. The methods are applied in the most ethical manner following good laboratory practice guidelines. The results of this report are based solely on the sample submitted and cannot be reproduced.